

 **ANALYTICAL REPORT****PREPARED FOR**

Attn: Brad Puetz
City of Sioux City
PO BOX 447
Sioux City, Iowa 51101
Generated 5/21/2025 5:38:12 PM

JOB DESCRIPTION

IA9778054 - 02/02 Southbridge

JOB NUMBER

810-147808-1

Eurofins Eaton Analytical South Bend

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Authorization



Generated
5/21/2025 5:38:12 PM

Authorized for release by
Traci Chlebowski, Senior Project Manager
Traci.Chlebowski@et.eurofinsus.com
(574)233-4777



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Client Sample Results	6
Lab Chronicle	9
Certification Summary	10
Method Summary	11
Sample Summary	12
Chain of Custody	13
Receipt Checklists	14

Definitions/Glossary

Client: City of Sioux City
Project/Site: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City of Sioux City
Project: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Job ID: 810-147808-1

Eurofins Eaton Analytical South Bend

Job Narrative 810-147808-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable.

- Matrix QC may not be reported if insufficient sample is provided or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 5/8/2025 9:15 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 3.0°C.

PFAS

Method 533: The pH of the following sample was adjusted to pH 7.5 in the laboratory: Finished Tap Southbridge (810-147808-1)

Method 533: The pH of the following sample was adjusted to pH 7.5 in the laboratory: FTB (810-147808-2)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: City of Sioux City
 Project/Site: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Client Sample ID: Finished Tap Southbridge

Lab Sample ID: 810-147808-1

Date Collected: 05/07/25 13:30

Matrix: Drinking Water

Date Received: 05/08/25 09:15

PWSID Number: IA9778054

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid	<0.78		1.9	0.78 ng/L		05/13/25 08:41	05/15/25 01:28	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.54		1.9	0.54 ng/L		05/13/25 08:41	05/15/25 01:28	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	1.7	J	1.9	0.63 ng/L		05/13/25 08:41	05/15/25 01:28	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	1.8	J	1.9	0.64 ng/L		05/13/25 08:41	05/15/25 01:28	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.57		1.9	0.57 ng/L		05/13/25 08:41	05/15/25 01:28	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	<0.92		1.9	0.92 ng/L		05/13/25 08:41	05/15/25 01:28	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	<0.67		1.9	0.67 ng/L		05/13/25 08:41	05/15/25 01:28	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.88		1.9	0.88 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.62		1.9	0.62 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.77		1.9	0.77 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.61		1.9	0.61 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorobutanesulfonic acid (PFBS)	20		1.9	0.62 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorobutanoic acid (PFBA)	9.7		1.9	0.49 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorodecanoic acid (PFDA)	<0.62		1.9	0.62 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorododecanoic acid (PFDoA)	<0.66		1.9	0.66 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.57		1.9	0.57 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluoroheptanoic acid (PFHpA)	4.5		1.9	0.68 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorohexanesulfonic acid (PFHxS)	35		1.9	0.62 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorohexanoic acid (PFHxA)	32		1.9	0.69 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorononanoic acid (PFNA)	<0.69		1.9	0.69 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorooctanesulfonic acid (PFOS)	4.8		1.9	0.65 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluorooctanoic acid (PFOA)	6.9		1.9	0.70 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluoropentanesulfonic acid (PFPeS)	13		1.9	0.65 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluoropentanoic acid (PFPeA)	21		1.9	0.73 ng/L		05/13/25 08:41	05/15/25 01:28	1
Perfluoroundecanoic acid (PFUnA)	<0.66		1.9	0.66 ng/L		05/13/25 08:41	05/15/25 01:28	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDoA	90		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C2-4:2-FTS	101		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C2-6:2-FTS	119		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C2-8:2-FTS	108		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C3 HFPO-DA	91		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C3 PFBS	103		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C3 PFHxS	100		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C4 PFBA	97		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C4 PFHpA	89		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C5 PFHxA	90		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C5 PFPeA	101		50 - 200	05/13/25 08:41	05/15/25 01:28	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: City of Sioux City
 Project/Site: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Client Sample ID: Finished Tap Southbridge

Lab Sample ID: 810-147808-1

Date Collected: 05/07/25 13:30

Matrix: Drinking Water

Date Received: 05/08/25 09:15

PWSID Number: IA9778054

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C6 PFDA	92		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C7 PFUnA	92		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C8 PFOA	93		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C8 PFOS	96		50 - 200	05/13/25 08:41	05/15/25 01:28	1
13C9 PFNA	90		50 - 200	05/13/25 08:41	05/15/25 01:28	1

Method: TAL SOP Total PFCA-Sum - Total PFCA (Summary)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Total PFOA and PFOS	12		2.0	0.20 ng/L			05/15/25 01:28	1

Client Sample ID: FTB

Lab Sample ID: 810-147808-2

Date Collected: 05/07/25 13:30

Matrix: Drinking Water

Date Received: 05/08/25 09:15

PWSID Number: IA9778054

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane e-1-sulfonic acid	<0.79		1.9	0.79 ng/L		05/19/25 10:21	05/20/25 20:37	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.55		1.9	0.55 ng/L		05/19/25 10:21	05/20/25 20:37	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.64		1.9	0.64 ng/L		05/19/25 10:21	05/20/25 20:37	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.65		1.9	0.65 ng/L		05/19/25 10:21	05/20/25 20:37	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.57		1.9	0.57 ng/L		05/19/25 10:21	05/20/25 20:37	1
9-Chlorohexadecafluoro-3-oxanonane e-1-sulfonic acid	<0.93		1.9	0.93 ng/L		05/19/25 10:21	05/20/25 20:37	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	<0.68		1.9	0.68 ng/L		05/19/25 10:21	05/20/25 20:37	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.89		1.9	0.89 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.63		1.9	0.63 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.78		1.9	0.78 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.62		1.9	0.62 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluorobutanesulfonic acid (PFBS)	<0.63		1.9	0.63 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluorobutanoic acid (PFBA)	<0.50		1.9	0.50 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluorodecanoic acid (PFDA)	<0.63		1.9	0.63 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluorododecanoic acid (PFDoA)	<0.67		1.9	0.67 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.57		1.9	0.57 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluoroheptanoic acid (PFHpA)	<0.69		1.9	0.69 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluorohexanesulfonic acid (PFHxS)	<0.63		1.9	0.63 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluoroheptanoic acid (PFHxA)	<0.70		1.9	0.70 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluorononanoic acid (PFNA)	<0.70		1.9	0.70 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluorooctanesulfonic acid (PFOS)	<0.66		1.9	0.66 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluorooctanoic acid (PFOA)	<0.71		1.9	0.71 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluoropentanesulfonic acid (PFPeS)	<0.66		1.9	0.66 ng/L		05/19/25 10:21	05/20/25 20:37	1
Perfluoropentanoic acid (PFPeA)	<0.74		1.9	0.74 ng/L		05/19/25 10:21	05/20/25 20:37	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: City of Sioux City
 Project/Site: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Client Sample ID: FTB

Lab Sample ID: 810-147808-2

Date Collected: 05/07/25 13:30

Matrix: Drinking Water

Date Received: 05/08/25 09:15

PWSID Number: IA9778054

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluoroundecanoic acid (PFUnA)	<0.67		1.9	0.67 ng/L		05/19/25 10:21	05/20/25 20:37	1
<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C2 PFDoA	89		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C2-4:2-FTS	91		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C2-6:2-FTS	90		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C2-8:2-FTS	109		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C3 HFPO-DA	91		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C3 PFBS	98		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C3 PFHxS	108		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C4 PFBA	95		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C4 PFHpA	95		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C5 PFHxA	97		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C5 PFPeA	103		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C6 PFDA	100		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C7 PFUnA	87		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C8 PFOA	92		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C8 PFOS	93		50 - 200			05/19/25 10:21	05/20/25 20:37	1
13C9 PFNA	97		50 - 200			05/19/25 10:21	05/20/25 20:37	1

Lab Chronicle

Client: City of Sioux City
 Project/Site: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Client Sample ID: Finished Tap Southbridge

Lab Sample ID: 810-147808-1

Date Collected: 05/07/25 13:30

Matrix: Drinking Water

Date Received: 05/08/25 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			143728	ET	EA SB	05/13/25 08:41
Total/NA	Analysis	533		1	144052	MH	EA SB	05/15/25 01:28
Total/NA	Analysis	Total PFCA-Sum		1	144318	CM	EA SB	05/15/25 01:28

Client Sample ID: FTB

Lab Sample ID: 810-147808-2

Date Collected: 05/07/25 13:30

Matrix: Drinking Water

Date Received: 05/08/25 09:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			144597	KB	EA SB	05/19/25 10:21
Total/NA	Analysis	533		1	144674	MH	EA SB	05/20/25 20:37

Laboratory References:

EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777



Accreditation/Certification Summary

Client: City of Sioux City
 Project/Site: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Laboratory: Eurofins Eaton Analytical South Bend

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
A2LA	ISO/IEC 17025	5794.01	07-31-26

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
Total PFCA-Sum		Drinking Water	Total PFOA and PFOS

Iowa	State	098	11-01-25
------	-------	-----	----------

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
Total PFCA-Sum		Drinking Water	Total PFOA and PFOS



Method Summary

Client: City of Sioux City
Project/Site: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA SB
Total PFCA-Sum	Total PFCA (Summary)	TAL SOP	EA SB
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA SB

Protocol References:

- EPA = US Environmental Protection Agency
- TAL SOP = TestAmerica Laboratories, Standard Operating Procedure

Laboratory References:

- EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777



Sample Summary

Client: City of Sioux City
Project/Site: IA9778054 - 02/02 Southbridge

Job ID: 810-147808-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
810-147808-1	Finished Tap Southbridge	Drinking Water	05/07/25 13:30	05/08/25 09:15	IA9778054
810-147808-2	FTB	Drinking Water	05/07/25 13:30	05/08/25 09:15	IA9778054

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11

Login Sample Receipt Checklist

Client: City of Sioux City

Job Number: 810-147808-1

Login Number: 147808

List Source: Eurofins Eaton Analytical South Bend

List Number: 1

Creator: Trowbridge, Peyton

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Were samples preserved to correct pH upon receipt, if applicable?	True	
Container provided by EEA	True	